

Product Specification

Chemical Oxygen Demand Standard, 100 ppm COD

Lot Number: SAMPLE

Product Number: 2227

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Acid Phthalate	877-24-7	ACS Acidimetric

Test	Specification	Result
Appearance	Colorless liquid	N/A
Chemical Oxygen Demand (COD)	0.0995-0.1005 mg/mL	N/A

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2227-16	500 mL natural poly	12 months
2227-32	1 L natural poly	12 months
2227-4	120 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)