Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

## **Product Specification**

Conductivity Standard, 23.8 µS/cm at 25°C (11.2 ppm TDS as NaCl)

Lot Number: SAMPLE Product Number: 2236.06

Manufacture Date: N/A Expiration Date: N/A

For use as a Total Dissolved Solids (TDS) standard: As Salt Water Ions 11.2 ppm as NaCl and 11.3 ppm as KCl; As Fresh Water Ions, 16 ppm as EPA/APHA, 15 ppm as "442".

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Chloride	7647-14-5	ACS

Appearance	Colorless liquid	Colorless liquid	*Not a certified value.
Test	Specification	Result	

Test	Certified Value	Uncertainty	NIST SRM#
Conductivity at 25°C (Method: SQCP031, SQCP033)	23.7-23.9 μS/cm	0.28 μS/cm	999

Conductivity measurements were performed in our Arlington, TX, Pocomoke City, MD, or Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
2236.06-1	4 L natural poly	12 months
2236.06-16	500 mL natural poly	12 months
2236.06-32	1 L natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.3 Lot Number: SAMPLE Product Number: 2236.06 Page 1 of 2

Version: 1.3	Lot Number: SAMPLE	Product Number: 2236.06	Page 2 of 2
This test report shall not	t be reproduced, except in full, without the written	n approval of Ricca Chemical Company.	
This product wa	s tested in an ISO 17025 Accredit	ted Laboratory	