

Product Specification

Nitric Acid, 10% (v/v)

Lot Number: SAMPLE

Product Number: 5320

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless to light yellow liquid	N/A	
Assay (vs. Sodium Hydroxide/Phenolphthalein)	9.9-10.1 % (v/v)	N/A	84

Specification	Reference
Nitric Acid	ASTM (D 287 M)
Nitric Acid (1+9)	ASTM (D 2907 A)
Nitric Acid (1+9)	ASTM (D 2972 C)
Nitric Acid (1+9)	ASTM (D 2187 H)
Nitric Acid (1+9)	ASTM (D 4548)
Nitric Acid (1+9)	ASTM (D 2187 I)
Nitric Acid (1+9)	ASTM (D 4638)
Nitric Acid (1+9)	ASTM (D 2187 J)
Nitric Acid (1+9)	ASTM (D 2187 K)
Nitric Acid (1+9)	ASTM (D 2187 L)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5320-1	4 L natural poly	24 months
5320-16	500 mL natural poly	24 months
5320-32	1 L natural poly	24 months
5320-5	20 L Cubitainer®	24 months
5320-55	55 gal poly drum	24 months
5320-5HP	20 L Ropak™	24 months
5320-5PT	2.5 L safety-coated glass	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.