Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

## **Product Specification**

Sulfuric Acid, 0.6667 Normal (2N/3) Suitable for Preparation of Protein-free Filtrates of Blood, Serum, or Plasma

Lot Number: SAMPLE Product Number: 8290

Manufacture Date: N/A Expiration Date: N/A

This item is used along with 10% Sodium Tungstate, RCC Cat. No. 7980.

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Sulfuric Acid	7664-93-9	ACS	

Test	Specification	$\mathbf{Result}$	NIST SRM#
Appearance	Colorless liquid	N/A	
Assay (vs.	0.6662-0.6672 N at 20°C	N/A	723
Tris(hydroxymethyl)aminomethane/Methyl Red)			

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
8290-1	4 L natural poly	36 months
8290-16	500 mL natural poly	36 months
8290-32	1 L natural poly	36 months
8290-5	20 L Cubitainer®	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: SAMPLE Product Number: 8290 Page 1 of 1