Arlington, TX 76012 Pocomoke City, MD 21851 Batesville, IN 47006 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

## **Product Specification**

## Turbidity Standard, 100 NTU, Polymer Bead Suspension

Lot Number: SAMPLE

Manufacture Date: N/A Expiration Date: N/A

Ready to use. Do NOT agitate. Pour gently into cuvette avoiding bubbles which will interfere with Turbidity readings.

Product Number: 8806

There is no NIST Standard Reference Material for Turbidity. This solution is traceable to commercially prepared EPA recognized Primary Standards

The EPA primary standard for turbidity (Formazin suspension, 4000 NTU) is identical to EP Primary Opalescent Suspension and USP Formazin Stock Suspension.

Due to the light scattering properties of the polystyrene Microspheres blend, your turbidity meter may respond differently and may require a specially formulated blend to perform within acceptable parameters. This 100 NTU Turbidity Standard is specially formulated to read 100 NTU on a Hach 2100N turbidity meter.

| Name                     | CAS#        | Grade           |
|--------------------------|-------------|-----------------|
| Water                    | 7732-18-5   | ACS/ASTM/USP/EP |
| Preservative             | Proprietary |                 |
| Polystyrene Microspheres | 9003-53-6   | Reagent         |

| Test       | Specification    | Result |
|------------|------------------|--------|
| Appearance | Colorless liquid | N/A    |
| Turbidity  | 99-101 NTU       | N/A    |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number | Size / Package Type | Shelf Life (Unopened Container) |
|-------------|---------------------|---------------------------------|
| 8806-1      | 4 L amber glass     | 12 months                       |
| 8806-16     | 500 mL clear PET-G  | 12 months                       |
| 8806-32     | 1 L amber glass     | 12 months                       |

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Page 1 of 1 Version: 1.3 Lot Number: SAMPLE Product Number: 8806