

Product Specification

Molybdenum ICP Standard, 1000 ppm Mo in 3% NH₄OH

Lot Number: SAMPLE

Product Number: PMO1KW

Manufacture Date: N/A

Expiration Date: N/A

This is a single element solution that was prepared volumetrically to contain the certified value reported. The uncertainty associated with the certified value is the sum of the estimated errors due to the purity of the raw material, the volumetric preparation of the solution, and transpiration of the solution through the container wall.

The final solution concentration is confirmed by AA, ICP, or ICP-MS, and is traceable to NIST Standard Reference Material 3134. All trace level elements were determined by ICP or ICP-MS.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ammonium Hydroxide	1336-21-6	Trace Metals
Ammonium Molybdate	13106-76-8	High Purity

Test	Specification	Result	NIST SRM#
Appearance	Colorless to slightly greenish-yellow liquid	N/A	
Molybdenum (Mo)	995-1005 ppm	N/A	3134

Trace Elements by ICP or ICP - MS

I=Spectral Interference N=Not Tested

All values reported in mg/L (ppm)

Aluminum (Al)	N/A	Iron (Fe)	N/A	Silver (Ag)	N/A
Antimony (Sb)	N/A	Lanthanum (La)	N/A	Sodium (Na)	N/A
Arsenic (As)	N/A	Lead (Pb)	N/A	Strontium (Sr)	N/A
Barium (Ba)	N/A	Lithium (Li)	N/A	Sulfur (S)	N/A
Beryllium (Be)	N/A	Lutetium (Lu)	N/A	Tantalum (Ta)	N/A
Bismuth (Bi)	N/A	Magnesium (Mg)	N/A	Tellurium (Te)	N/A
Boron (B)	N/A	Manganese (Mn)	N/A	Terbium (Tb)	N/A
Cadmium (Cd)	N/A	Mercury (Hg)	N/A	Thallium (Tl)	N/A
Calcium (Ca)	N/A	Neodymium (Nd)	N/A	Thorium (Th)	N/A
Cerium (Ce)	N/A	Nickel (Ni)	N/A	Thulium (Tm)	N/A
Cesium (Cs)	N/A	Niobium (Nb)	N/A	Tin (Sn)	N/A
Chromium (Cr)	N/A	Osmium (Os)	N/A	Titanium (Ti)	N/A
Cobalt (Co)	N/A	Palladium (Pd)	N/A	Tungsten (W)	N/A
Copper (Cu)	N/A	Phosphorus (P)	N/A	Uranium (U)	N/A
Dysprosium (Dy)	N/A	Platinum (Pt)	N/A	Vanadium (V)	N/A
Erbium (Er)	N/A	Potassium (K)	N/A	Ytterbium (Yb)	N/A
Europium (Eu)	N/A	Praseodymium (Pr)	N/A	Yttrium (Y)	N/A
Gadolinium (Gd)	N/A	Rhenium (Re)	N/A	Zinc (Zn)	N/A
Gallium (Ga)	N/A	Rhodium (Rh)	N/A	Zirconium (Zr)	N/A
Germanium (Ge)	N/A	Rubidium (Rb)	N/A		
Gold (Au)	N/A	Ruthenium (Ru)	N/A		
Hafnium (Hf)	N/A	Samarium (Sm)	N/A		
Holmium (Ho)	N/A	Scandium (Sc)	N/A		
Indium (In)	N/A	Selenium (Se)	N/A		
Iridium (Ir)	N/A	Silicon (Si)	N/A		

Specification	Reference
---------------	-----------

Molybdenum ICP Standard, 1 mL = 1 mg Mo (1,000 ppm Mo)
(NH₄)₂MoO₄ in H₂O / trace NH₄OH

EPA (200.7)

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
-------------	---------------------	---------------------------------

PMO1KW-100 100 mL natural LDPE 18 months

PMO1KW-500 500 mL natural poly 18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.