

Product Specification

Antimony ICP Standard, 1000 ppm Sb in H₂O/Tartaric Acid/tr HNO₃

Lot Number: SAMPLE

Product Number: PSB1KW

Manufacture Date: N/A

Expiration Date: N/A

This is a single element solution that was prepared volumetrically to contain the certified value reported. The uncertainty associated with the certified value is the sum of the estimated errors due to the purity of the raw material, the volumetric preparation of the solution, and transpiration of the solution through the container wall.

The final solution concentration is confirmed by AA, ICP, or ICP-MS, and is traceable to NIST Standard Reference Material 3102. All trace level elements were determined by ICP or ICP-MS.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Tartaric Acid	87-69-4	ACS
Nitric Acid	7697-37-2	Trace Metals
Hydrochloric Acid	7647-01-0	Trace Metals
Antimony	7440-36-0	High Purity

Test	Specification	Result	NIST SRM#
Antimony (Sb)	995-1005 ppm	N/A	3102
Appearance	Colorless liquid	N/A	

Trace Elements by ICP or ICP - MS

I=Spectral Interference N=Not Tested

All values reported in mg/L (ppm)

Aluminum (Al)	N/A	Iron (Fe)	N/A	Selenium (Se)	N/A
Arsenic (As)	N/A	Lanthanum (La)	N/A	Silicon (Si)	N/A
Barium (Ba)	N/A	Lead (Pb)	N/A	Silver (Ag)	N/A
Beryllium (Be)	N/A	Lithium (Li)	N/A	Sodium (Na)	N/A
Bismuth (Bi)	N/A	Lutetium (Lu)	N/A	Strontium (Sr)	N/A
Boron (B)	N/A	Magnesium (Mg)	N/A	Sulfur (S)	N/A
Cadmium (Cd)	N/A	Manganese (Mn)	N/A	Tantalum (Ta)	N/A
Calcium (Ca)	N/A	Mercury (Hg)	N/A	Tellurium (Te)	N/A
Cerium (Ce)	N/A	Molybdenum (Mo)	N/A	Terbium (Tb)	N/A
Cesium (Cs)	N/A	Neodymium (Nd)	N/A	Thallium (Tl)	N/A
Chromium (Cr)	N/A	Nickel (Ni)	N/A	Thorium (Th)	N/A
Cobalt (Co)	N/A	Niobium (Nb)	N/A	Thulium (Tm)	N/A
Copper (Cu)	N/A	Osmium (Os)	N/A	Tin (Sn)	N/A
Dysprosium (Dy)	N/A	Palladium (Pd)	N/A	Titanium (Ti)	N/A
Erbium (Er)	N/A	Phosphorus (P)	N/A	Tungsten (W)	N/A
Europium (Eu)	N/A	Platinum (Pt)	N/A	Uranium (U)	N/A
Gadolinium (Gd)	N/A	Potassium (K)	N/A	Vanadium (V)	N/A
Gallium (Ga)	N/A	Praseodymium (Pr)	N/A	Ytterbium (Yb)	N/A
Germanium (Ge)	N/A	Rhenium (Re)	N/A	Yttrium (Y)	N/A
Gold (Au)	N/A	Rhodium (Rh)	N/A	Zinc (Zn)	N/A
Hafnium (Hf)	N/A	Rubidium (Rb)	N/A	Zirconium (Zr)	N/A
Holmium (Ho)	N/A	Ruthenium (Ru)	N/A		
Indium (In)	N/A	Samarium (Sm)	N/A		
Iridium (Ir)	N/A	Scandium (Sc)	N/A		

Specification	Reference
---------------	-----------

Antimony ICP Standard, 1000 ppm

EPA (200.7)

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
-------------	---------------------	---------------------------------

PSB1KW-100

100 mL natural LDPE

18 months

PSB1KW-500

500 mL natural poly

18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.