

# Product Specification

## Perchloric Acid, 0.01 Normal (N/100) in Dioxane

**Lot Number:** SAMPLE**Product Number:** R5513500**Manufacture Date:** N/A**Expiration Date:** N/A

| Name            | CAS#      | Grade |
|-----------------|-----------|-------|
| Perchloric Acid | 7601-90-3 | ACS   |
| Water           | 7732-18-5 | ACS   |
| Dioxane         | 123-91-1  | HPLC  |

| Test                                                    | Specification           | Result | NIST SRM# |
|---------------------------------------------------------|-------------------------|--------|-----------|
| Appearance                                              | Colorless liquid        | N/A    |           |
| Assay (vs. Potassium Hydrogen Phthalate/Crystal Violet) | 0.0095-0.0105 N at 20°C | N/A    | 84        |

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

| Part Number   | Size / Package Type | Shelf Life (Unopened Container) |
|---------------|---------------------|---------------------------------|
| R5513500-1C   | 1 L amber glass     | 12 months                       |
| R5513500-500C | 500 mL amber glass  | 12 months                       |

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)