

Product Specification

Tightrant™ Potassium Hydroxide, 0.100 Normal (N/10)
Tested in an ISO 17025 Accredited Facility

Lot Number: SAMPLE

Product Number: 6200

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Hydroxide	1310-58-3	ACS

Test	Specification	Result
Appearance	Colorless liquid	Colorless liquid *Not a certified value.

Test	Certified Value	Uncertainty	NIST SRM#
Assay (vs. Potassium Hydrogen Phthalate/Phenolphthalein) (Method: SQCP008, SQCP033)	0.0999-0.1001 N at 20°C	0.0002	84

Specification	Reference
Potassium Hydroxide, 0.1 M	EPA (SW-846) (61)
Potassium Hydroxide, 0.1 N aqueous solution	ASTM (D 2619)
Potassium hydroxide, 0.1 N	TAPPI (T 610 sp-97)
Potassium hydroxide, 0.1 N	TAPPI (T 689 om-93)
0.1 M Potassium hydroxide Volumetric Solution	EP (3004800)

Titration measurements were performed in our Arlington, TX, Pocomoke City, MD, or Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6200-1	4 L natural poly	24 months
6200-16	500 mL natural poly	24 months
6200-1CT	4 L Cubitainer®	24 months
6200-2.5	10 L Cubitainer®	24 months
6200-32	1 L natural poly	24 months
6200-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.