Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Tightrant™ Potassium Hydroxide, 1.00 Normal Tested in an ISO 17025 Accredited Facility

Lot Number: SAMPLE Product Number: 6250

Manufacture Date: N/A

Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Hydroxide	1310-58-3	ACS

Test	Specification	Result	
Appearance	Colorless liquid	Colorless liquid	d *Not a certified value
Test	Certified Value	Uncertainty	NIST SRM#
Assay (vs. Potassium Hydrogen Phthalate/Phenolphthalein) (Method: SQCP008,	0.999-1.001 N at 20°C	0.002	84

Specification	Reference
Potassium hydroxide, 1.0 N	TAPPI (T 610 sp-97)
1 M Potassium hydroxide Volumetric Solution	EP (3009100)
Potassium Hydroxide, Normal (1 N) Volumetric Solution (modified)	USP (N/A)

Titration measurements were performed in our Arlington, TX, Pocomoke City, MD, or Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
6250-1	4 L natural poly	24 months
6250-16	500 mL natural poly	24 months
6250-1CT	4 L Cubitainer®	24 months
6250-2.5	10 L Cubitainer®	24 months
6250-32	1 L natural poly	24 months
6250-5	20 L Cubitainer®	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.4 Lot Number: SAMPLE Product Number: 6250 Page 1 of 2

7	Let Number: SAMPLE	Product Number: 6250	Page 2 of 2
This test report shall not be repr	oduced, except in full, without the writte	en approval of Ricca Chemical Company.	
This product was test	ed in an ISO 17025 Accredi	ted Laboratory	

Version: 1.4 Lot Number: SAMPLE Product Number: 6250 Page 2 of