

## **Product Specification**

## Sulfuric Acid, 49% (w/w)

Lot Number: SAMPLE	Product Number: R8179490	Manufacture Date: N/A
		Expiration Date: N/A

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Sulfuric Acid	7664-93-9	ACS	
Test	Specification	Result	NIST SRM#

L		Specification	Result	
Α	ppearance	Colorless liquid	N/A	
А	ssay (vs. Sodium Hydroxide/Phenolphthalein)	48.5-49.5 % (w/w)	N/A	84

Specification	Reference
Sulfuric acid, 1:1	TAPPI (T 618 cm-84)
Sulfuric acid, 1:1	TAPPI (T 632 cm-84)
Sulfuric Acid 1+1	EPA (377.1)
Sulfuric Acid/Water (1:1, v/v)	EPA (SW-846) (3665A)
Sulfuric Acid (1:1 v/v)	EPA (SW-846) (3650B)
Sulfuric Acid Solution, 1 + 1	EPA (376.2)
Sulfuric Acid Solution (1:1 v/v)	EPA (SW-846) (3535)
Sulfuric Acid Solution (1:1 v/v)	EPA (SW-846) (3510C)
Sulfuric Acid Solution (1:1 v/v)	EPA (SW-846) (3520C)
Sulfuric Acid (9 Molar)	EPA (SW-846) (3542)
Sampler Wash Solution	EPA (410.4)
Sulfuric acid, 1+1	TAPPI (T 665 om-93)
Sulfuric Acid (1+1)	ASTM (D 874)
Sulfuric Acid (1+1)	·····

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R8179490-10F	10 L Cubitainer®	36 months
D 110		

Recommended Storage:  $15^{\circ}C - 30^{\circ}C (59^{\circ}F - 86^{\circ}F)$ 

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.