Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

High Level Standard

Lot Number: SAMPLE Product Number: RCELHIGH

Manufacture Date: N/A Expiration Date: N/A

Matrix: 2% Nitric Acid / 0.05% HCl

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals
Hydrochloric Acid	7647-01-0	Trace Metals

Test	Specification	Result
Appearance	Light yellow liquid	N/A

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Aluminum (Al)	455.4-464.6 ppm	Aluminum Nitrate Nonahydrate	7784-27-2	High Purity	3101
Iron (Fe)	455.4-464.6 ppm	Iron	7439-89-6	High Purity	3126
Chromium (Cr)	9.9-10.1	Chromium Nitrate Nonahydrate	7789-02-8	High Purity	3112
Selenium (Se)	9.9-10.1 ppm	Selenium	7782-49-2	High Purity	3149
Zinc (Zn)	9.9-10.1 ppm	Zinc	7440-66-6	High Purity	3168
Copper (Cu)	9.9-10.1 ppm	Copper	7440-50-8	High Purity	3114
Nickel (Ni)	9.9-10.1 ppm	Nickel	7440-02-0	High Purity	3136

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RCELHIGH-500N	500 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Version: 1.0 Lot Number: SAMPLE Product Number: RCELHIGH Page 1 of 2

Version: 1.0	Lot Number: SAMPLE	Product Number: RCELHIGH	Page 2 of 2
This test report shall not be rep	roduced, except in full, without the writte	en approval of Ricca Chemical Company.	