Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

## **Product Specification**

Mixed Anion Standard 1 ppm F, 10 ppm Cl, 10 ppm SO<sub>4</sub><sup>2</sup>, 10 ppm Br

Lot Number: SAMPLE Product Number: ICMIX03

Manufacture Date: N/A Expiration Date: N/A

The certified values reported are the prepared values based upon the method of preparation of the material. The uncertainties in the prepared values are based upon the volumetric method of preparation.

Matrix: Water

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP

Test	Specification	Result
Appearance	Colorless liquid	N/A

Analyte	Analysis (ppm)	Solute	CAS#	Grade
Chloride (Cl)	9.95-10.05 ppm	Sodium Chloride	7647-14-5	ACS
Sulfate (SO4)	9.95-10.05 ppm	Sodium Sulfate Anhydrous	7757-82-6	ACS
Bromide (Br)	9.95-10.05 ppm	Sodium Bromide	7647-15-6	High Purity
Fluoride (F)	0.995-1.005 ppm	Sodium Fluoride	7681-49-4	High Purity

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
ICMIX03-100	100 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.4 Lot Number: SAMPLE Product Number: ICMIX03 Page 1 of 1