Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Mixed Calibration Standard 3, ICP, EPA 200.7 version 4.4

Lot Number: SAMPLE Product Number: RMCAL3

Manufacture Date: N/A
Expiration Date: N/A

This is a multielement solution that was prepared volumetrically to contain the elements and concentrations stated below. The uncertainty associated with the certified values is $\pm 1\%$ relative. The concentrations are confirmed by ICP or ICP-MS.

Matrix: 5% Nitric Acid

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Nitric Acid	7697-37-2	Trace Metals	

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Phosphorus (P)	99-101 ppm	Ammonium Dihydrogen Phosphate	7722-76-1	High Purity	3139
Cerium (Ce)	19.8-20.2 ppm	Cerium (III) Nitrate Hexahydrate	10294-41-4	High Purity	3110
Vanadium (V)	19.8-20.2 ppm	Vanadium	7440-62-2	High Purity	3165
Cobalt (Co)	19.8-20.2 ppm	Cobalt	7440-48-4	High Purity	3113

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)	
RMCAL3-100	100 mL natural LDPE	12 months	

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: SAMPLE Product Number: RMCAL3 Page 1 of 1