Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

ICP-MS Tuning Solution, Method 6020 (91)

Lot Number: SAMPLE Product Number: RMSTUNE1

Manufacture Date: N/A Expiration Date: N/A

stated below. The uncertainty

This is a multielement solution that was prepared volumetrically to contain the elements and concentrations stated below. The uncertainty associated with the certified values is $\pm 0.5\%$ relative. The concentrations are confirmed by ICP or ICP-MS.

Matrix: 2% Nitric Acid

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals
Lithium Carbonate, 6Li2CO3	554-13-2	

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	N/A	
Lithium (Li)	9.95-10.05 ppm	N/A	3129

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Indium (In)	9.95-10.05 ppm	Indium	7440-74-6	High Purity	3124
Cobalt (Co)	9.95-10.05 ppm	Cobalt	7440-48-4	High Purity	3113
Thallium (Tl)	9.95-10.05 ppm	Thallium	7440-28-0	High Purity	3158

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RMSTUNE1-100	100 mL natural LDPE	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: SAMPLE Product Number: RMSTUNE1 Page 1 of 1