

Product Specification

Mixed Calibration Standard 4, ICP

Lot Number: SAMPLE**Product Number:** RMXCL4**Manufacture Date:** N/A**Expiration Date:** N/A

This is a multielement solution that was prepared gravimetrically to contain the elements and concentrations stated below. The uncertainty associated with the certified values is $\pm 0.5\%$ relative. The concentrations are confirmed by ICP or ICP-MS.

Matrix: 2% Nitric Acid

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals

Test	Specification	Result
Appearance	Colorless liquid	N/A

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Aluminum (Al)	199-201 ppm	Aluminum Nitrate Nonahydrate	7784-27-2	High Purity	3101
Calcium (Ca)	995-1005 ppm	Calcium Carbonate	471-34-1	High Purity	3109
Potassium (K)	398-402 ppm	Potassium Carbonate	584-08-7	High Purity	3141
Sodium (Na)	199-201 ppm	Sodium Carbonate	497-19-8	High Purity	3152
Chromium (Cr)	19.9-20.1 ppm	Chromium Nitrate Nonahydrate	7789-02-8	High Purity	3112
Nickel (Ni)	19.9-20.1 ppm	Nickel	7440-02-0	High Purity	3136

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RMXCL4-100	100 mL natural LDPE	12 months
RMXCL4-500	500 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.