

Product Specification

Mixed Calibration Standard 5R, ICP

Lot Number: SAMPLE**Product Number:** RMXCL5R**Manufacture Date:** N/A**Expiration Date:** N/A

This is a multielement solution that was prepared gravimetrically to contain the elements and concentrations stated below. The uncertainty associated with the certified values is $\pm 0.5\%$ relative. The concentrations are confirmed by ICP or ICP-MS.

Matrix: 2% Nitric Acid

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals
Tartaric Acid	87-69-4	ACS
Hydrochloric Acid	7647-01-0	Trace Metals

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Magnesium (Mg)	995-1005 ppm	Magnesium Chloride	7786-30-3	High Purity	3131
Antimony (Sb)	198-202 ppm	Antimony	7440-36-0	High Purity	3102
Thallium (Tl)	198-202 ppm	Thallium	7440-28-0	High Purity	3158
Silver (Ag)	49.7-50.3 ppm	Silver Nitrate	7761-88-8	High Purity	3151

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RMXCL5R-500	500 mL amber poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.