

Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

## **Product Specification**

Chromium ICP Standard, 100 ppm Cr in dilute HNO<sub>3</sub>

Lot Number: SAMPLE Product Number: RPCR01KN

Manufacture Date: N/A
Expiration Date: N/A

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals
Chromium Nitrate Nonahydrate	7789-02-8	High Purity

Test	Specification	Result	NIST SRM#
Appearance	Colorless to light blue liquid	N/A	
Chromium (Cr)	90-110 ppm	N/A	3112

Specification	Reference
Chromium ICP, 1,000 ppm in HNO3	EPA (200.7)

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RPCR01KN-100N	100 mL natural LDPE	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: SAMPLE Product Number: RPCR01KN Page 1 of 1