

Product Specification

Mixed Anion Standard, 10 ppm: Br, Cl

Lot Number: SAMPLE

Product Number: RPMX040N

Manufacture Date: N/A

Recommended Retest Date: N/A

Matrix: Water

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP

Test	Specification	Result
Appearance	Colorless liquid	N/A

Analyte	Analysis (ppm)	Solute	CAS#	Grade
Chloride (Cl)	9.9-10.1 ppm	Sodium Chloride	7647-14-5	ACS
Bromide (Br)	9.9-10.1 ppm	Sodium Bromide	7647-15-6	High Purity

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RPMX040N-100N	100 mL natural LDPE	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)