Arlington, TX 76012
Pocomoke City, MD 21851
Batesville, IN 47006
http://www.riccachemical.com
1-888-GO-RICCA
customerservice@riccachemical.com

Product Specification

Spiking Standard #2R

Lot Number: SAMPLE Product Number: RSPK2RDW

Manufacture Date: N/A

Expiration Date: N/A

This is a multielement solution that was prepared gravimetrically to contain the elements and concentrations stated below. The uncertainty associated with the certified values is $\pm 0.5\%$ relative. The concentrations are confirmed by ICP or ICP-MS.

Matrix: 2% Nitric Acid

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Nitric Acid	7697-37-2	Trace Metals

Test	Specification	Result
Appearance	Colorless liquid	N/A

Analyte	Analysis (ppm)	Solute	CAS#	Grade	NIST SRM#
Calcium (Ca)	9950-10050 ppm	Calcium Carbonate	471-34-1	High Purity	3109
Sodium (Na)	9950-10050 ppm	Sodium Carbonate	497-19-8	High Purity	3152
Potassium (K)	9950-10050 ppm	Potassium Carbonate	584-08-7	High Purity	3141
Magnesium (Mg)	9950-10050 ppm	Magnesium	7439-95-4	High Purity	3131

This standard is guaranteed to be stable and accurate provided the product is kept tightly capped and stored under normal laboratory conditions. Balances are calibrated using NIST traceable weights whose verification of maintenance and recalibration is documented per in-house Standard Operating Procedures. Class A glassware is also calibrated and routinely rechecked per in-house Standard Operating Procedures. Trace metal analyzed acids and Trace Metals Analyzed Water are used in the manufacture of this product. Triple cleaned containers are used in the manufacture of this product.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
RSPK2RDW-50N	50 mL natural poly	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.2 Lot Number: SAMPLE Product Number: RSPK2RDW Page 1 of 1